

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE		
Calibration File No.:	01531	Calib. Date:	10/12/2017
Certification File No.:	01532	Cert. No.:	00034
Linearity File No.:	01533	Cert. Date:	10/12/2017
Solution File No.:	01565	Lin. No.:	00025
Sequential File No.:	01565	Lin. Date:	10/12/2017
		Soln. No.:	00268
		Soln. Date:	01/01/2018
		File Date:	01/01/2018

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWL S3-0440
Control Solution %:	0.100%			Expires:	12/02/2018
Solution Control Lot:	16390			Bottle No.:	1199

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:29S		
Control 1 EC	0.095%	03:30S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	03:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:31S		
Control 2 EC	0.095%	03:32S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	03:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:32S		
Control 3 EC	0.095%	03:33S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	03:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: _____

Badge No.: 128

Date: 01/01/2018

