

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01531 **Calib. Date:** 10/12/2017 **Calib. No.:** 00034
Certification File No.: 01532 **Cert. Date:** 10/12/2017 **Cert. No.:** 00025
Linearity File No.: 01533 **Lin. Date:** 10/12/2017 **Lin. No.:** 00025
Solution File No.: 01567 **Soln. Date:** 01/01/2018 **Soln. No.:** 00269
Sequential File No.: 01567 **File Date:** 01/01/2018

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWL S3-0440
Control Solution %: 0.100% **Expires:** 12/02/2018
Solution Control Lot: 16390 **Bottle No.:** 1306

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:40S		
Control 1 EC	0.098%	23:41S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:42S		
Control 2 EC	0.096%	23:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:43S		
Control 3 EC	0.098%	23:44S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PASTEKA

First Name: EDWARD

MI: -

Signature: 

Badge No.: 123

Date: 01/01/2018