

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01531 Calib. Date: 10/12/2017 Calib. No.: 00034
Certification File No.: 01532 Cert. Date: 10/12/2017 Cert. No.: 00025
Linearity File No.: 01533 Lin. Date: 10/12/2017 Lin. No.: 00025
Solution File No.: 01574 Soln. Date: 01/20/2018 Soln. No.: 00270
Sequential File No.: 01574 File Date: 01/20/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0440
Control Solution %: 0.100% Expires: 12/02/2018
Solution Control Lot: 16390 Bottle No.: 1223

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:42S		
Control 1 EC	0.098%	02:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:43S		
Control 2 EC	0.096%	02:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.095%	02:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:45S		
Control 3 EC	0.095%	02:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.095%	02:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:46S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: 

Badge No.: 120

Date: 01/20/2018