

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01583	Calib. Date: 02/16/2018
Certification File No.:	01584	Calib. No.: 00035
Linearity File No.:	01585	Cert. Date: 02/16/2018
Solution File No.:	01596	Cert. No.: 00026
Sequential File No.:	01596	Lin. Date: 02/16/2018
		Lin. No.: 00026
		Soln. Date: 03/10/2018
		Soln. No.: 00273
		File Date: 03/10/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0440
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0042

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:40S		
Control 1 EC	0.094%	22:41S	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.096%	22:41S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	22:42S		
Control 2 EC	0.095%	22:42S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	22:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:43S		
Control 3 EC	0.096%	22:44S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:44S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: _____

Badge No.: 128

Date: 03/10/2018

