

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01583 Calib. Date: 02/16/2018 Calib. No.: 00035
Certification File No.: 01584 Cert. Date: 02/16/2018 Cert. No.: 00026
Linearity File No.: 01585 Lin. Date: 02/16/2018 Lin. No.: 00026
Solution File No.: 01610 Soln. Date: 05/10/2018 Soln. No.: 00276
Sequential File No.: 01610 File Date: 05/10/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0440
Control Solution %: 0.100% Expires: 12/11/2019
Solution Control Lot: 17420 Bottle No.: 0257

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:30D		
Control 1 EC	0.101%	22:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:32D		
Control 2 EC	0.099%	22:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:33D		
Control 3 EC	0.100%	22:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 05/10/2018