

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01583	Calib. Date: 02/16/2018
Certification File No.:	01584	Calib. No.: 00035
Linearity File No.:	01585	Cert. Date: 02/16/2018
Solution File No.:	01634	Cert. No.: 00026
Sequential File No.:	01634	Lin. Date: 02/16/2018
		Lin. No.: 00026
		Soln. Date: 07/07/2018
		Soln. No.: 00279
		File Date: 07/07/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0440
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0249

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:36D		
Control 1 EC	0.101%	01:36D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	01:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:37D		
Control 2 EC	0.099%	01:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:38D		
Control 3 EC	0.099%	01:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: *Shawn Heckler #120*

Badge No.: 120

Date: 07/07/2018