

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	Calib. No.:	00038
Calibration File No.:	01689	Cert. No.:	00029
Certification File No.:	01690	Lin. No.:	00029
Linearity File No.:	01691	Soln. No.:	00289
Solution File No.:	01700	Calib. Date:	01/15/2019
Sequential File No.:	01700	Cert. Date:	01/15/2019
		Lin. Date:	01/15/2019
		Soln. Date:	02/14/2019
		File Date:	02/14/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWL S3-0440
Control Solution %:	0.100%			Expires:	12/11/2019
Solution Control Lot:	17420			Bottle No.:	0336

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:20S		
Control 1 EC	0.099%	18:20S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:21S		
Control 2 EC	0.097%	18:22S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:22S		
Control 3 EC	0.098%	18:23S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 02/14/2019