

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01689	Calib. Date: 01/15/2019
Certification File No.:	01690	Calib. No.: 00038
Linearity File No.:	01691	Cert. Date: 01/15/2019
Solution File No.:	01719	Cert. No.: 00029
Sequential File No.:	01719	Lin. Date: 01/15/2019
		Lin. No.: 00029
		Soln. Date: 04/24/2019
		Soln. No.: 00292
		File Date: 04/24/2019

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0440
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0332

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:50D		
Control 1 EC	0.099%	23:51D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:51D		
Control 2 EC	0.099%	23:52D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:53D		
Control 3 EC	0.099%	23:54D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:54D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: 

Badge No.: 120

Date: 04/24/2019