

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01689 Calib. Date: 01/15/2019 Calib. No.: 00038
Certification File No.: 01690 Cert. Date: 01/15/2019 Cert. No.: 00029
Linearity File No.: 01691 Lin. Date: 01/15/2019 Lin. No.: 00029
Solution File No.: 01727 Soln. Date: 05/24/2019 Soln. No.: 00293
Sequential File No.: 01727 File Date: 05/24/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0440
Control Solution %: 0.100% Expires: 12/11/2019
Solution Control Lot: 17420 Bottle No.: 0337

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:41D		
Control 1 EC	0.099%	23:42D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:43D		
Control 2 EC	0.098%	23:43D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:44D		
Control 3 EC	0.099%	23:45D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:46D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 05/24/2019