

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01739 Calib. Date: 06/21/2019 Calib. No.: 00039
Certification File No.: 01740 Cert. Date: 06/21/2019 Cert. No.: 00030
Linearity File No.: 01741 Lin. Date: 06/21/2019 Lin. No.: 00030
Solution File No.: 01746 Soln. Date: 07/05/2019 Soln. No.: 00295
Sequential File No.: 01746 File Date: 07/05/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0441
Control Solution %: 0.100% Expires: 01/08/2021
Solution Control Lot: 19020 Bottle No.: 1316

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:57D		
Control 1 EC	0.099%	01:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:58D		
Control 2 EC	0.099%	01:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:59D		
Control 3 EC	0.099%	02:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 07/05/2019