

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWJ-0018  
Location: SEASIDE HEIGHTS POLICE  
Calibration File No.: 01739 Calib. Date: 06/21/2019 Calib. No.: 00039  
Certification File No.: 01740 Cert. Date: 06/21/2019 Cert. No.: 00030  
Linearity File No.: 01741 Lin. Date: 06/21/2019 Lin. No.: 00030  
Solution File No.: 01767 Soln. Date: 10/03/2019 Soln. No.: 00298  
Sequential File No.: 01767 File Date: 10/03/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0441  
Control Solution %: 0.100% Expires: 01/08/2021  
Solution Control Lot: 19020 Bottle No.: 1140

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:03D		
Control 1 EC	0.100%	16:04D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:05D		
Control 2 EC	0.098%	16:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:06D		
Control 3 EC	0.098%	16:07D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:08D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 10/03/2019