

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE		
Calibration File No.:	01775	Calib. Date:	11/27/2019
Certification File No.:	01776	Cert. Date:	11/27/2019
Linearity File No.:	01777	Lin. Date:	11/27/2019
Solution File No.:	01791	Soln. Date:	02/14/2020
Sequential File No.:	01791	File Date:	02/14/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWL S3-0440
Solution Control Lot:	19020	Expires:	01/08/2021
		Bottle No.:	1416

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:03S		
Control 1 EC	0.099%	04:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	04:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:05S		
Control 2 EC	0.096%	04:06S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	04:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:06S		
Control 3 EC	0.097%	04:07S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: *Shawn Heckler #120*

Badge No.: 120

Date: 02/14/2020