

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
 Location: SEASIDE HEIGHTS POLICE
 Serial No.: ARWJ-0018
 Calibration File No.: 01848
 Certification File No.: 01849
 Linearity File No.: 01850
 Solution File No.: 01862
 Sequential File No.: 01862
 Calib. Date: 09/29/2020
 Cert. Date: 09/29/2020
 Lin. Date: 09/29/2020
 Soln. Date: 12/01/2020
 File Date: 12/01/2020
 Calib. No.: 00042
 Cert. No.: 00033
 Lin. No.: 00033
 Soln. No.: 00323

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 20320
 Model No.: CU-34
 Serial No.: DDWL S3-0440
 Expires: 07/06/2022
 Bottle No.: 0578

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:35S		
Control 1 EC	0.101%	16:36S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:36S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:37S		
Control 2 EC	0.098%	16:37S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:37S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:38S		
Control 3 EC	0.098%	16:39S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:39S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 12/01/2020