

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWJ-0018  
Location: SEASIDE HEIGHTS POLICE  
Calibration File No.: 01875 Calib. Date: 03/16/2021 Calib. No.: 00043  
Certification File No.: 01876 Cert. Date: 03/16/2021 Cert. No.: 00034  
Linearity File No.: 01877 Lin. Date: 03/16/2021 Lin. No.: 00034  
Solution File No.: 01904 Soln. Date: 06/29/2021 Soln. No.: 00331  
Sequential File No.: 01904 File Date: 06/29/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0441  
Control Solution %: 0.100% Expires: 01/13/2023  
Solution Control Lot: 21010 Bottle No.: 0497

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:13D		
Control 1 EC	0.099%	20:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	20:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:14D		
Control 2 EC	0.098%	20:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:16D		
Control 3 EC	0.098%	20:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:17D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 06/29/2021