

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE		
Calibration File No.:	01875	Calib. Date:	03/16/2021
Certification File No.:	01876	Cert. No.:	00043
Linearity File No.:	01877	Cert. Date:	03/16/2021
Solution File No.:	01918	Lin. No.:	00034
Sequential File No.:	01918	Lin. Date:	03/16/2021
		Soln. Date:	07/18/2021
		Soln. No.:	00332
		File Date:	07/18/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWL S3-0441
Solution Control Lot:	21010	Expires:	01/13/2023
		Bottle No.:	0212

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:09D		
Control 1 EC	0.099%	10:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10D		
Control 2 EC	0.099%	10:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:12D		
Control 3 EC	0.099%	10:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 07/18/2021