

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE		
Calibration File No.:	01923	Calib. Date:	07/27/2021
Certification File No.:	01924	Cert. No.:	00044
Linearity File No.:	01925	Cert. Date:	07/27/2021
Solution File No.:	01933	Lin. No.:	00035
Sequential File No.:	01933	Lin. Date:	07/27/2021
		Soln. Date:	08/24/2021
		Soln. No.:	00334
		File Date:	08/24/2021

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWL S3-0440
Control Solution %:	0.100%			Expires:	01/13/2023
Solution Control Lot:	21010			Bottle No.:	0152

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:28D		
Control 1 EC	0.099%	18:29D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:29D		
Control 2 EC	0.099%	18:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:31D		
Control 3 EC	0.099%	18:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:32D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 08/24/2021