

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01923	Calib. Date: 07/27/2021
Certification File No.:	01924	Cert. No.: 00044
Linearity File No.:	01925	Cert. Date: 07/27/2021
Solution File No.:	01963	Lin. No.: 00035
Sequential File No.:	01963	Lin. Date: 07/27/2021
		Soln. No.: 00340
		Soln. Date: 12/03/2021
		File Date: 12/03/2021

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWL S3-0440
Control Solution %: 0.100%		Expires: 01/13/2023
Solution Control Lot: 21010		Bottle No.: 0050

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:29S		
Control 1 EC	0.100%	21:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:30S		
Control 2 EC	0.097%	21:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:32S		
Control 3 EC	0.098%	21:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:33S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 12/03/2021