

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWJ-0018  
Location: SEASIDE HEIGHTS POLICE  
Calibration File No.: 01923 Calib. Date: 07/27/2021 Calib. No.: 00044  
Certification File No.: 01924 Cert. Date: 07/27/2021 Cert. No.: 00035  
Linearity File No.: 01925 Lin. Date: 07/27/2021 Lin. No.: 00035  
Solution File No.: 01973 Soln. Date: 01/01/2022 Soln. No.: 00341  
Sequential File No.: 01973 File Date: 01/01/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0440  
Control Solution %: 0.100% Expires: 01/13/2023  
Solution Control Lot: 21010 Bottle No.: 0155

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:25S		
Control 1 EC	0.098%	02:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:27S		
Control 2 EC	0.098%	02:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:28S		
Control 3 EC	0.098%	02:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 01/01/2022