

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE		
Calibration File No.:	01975	Calib. Date:	01/20/2022
Certification File No.:	01976	Cert. No.:	00045
Linearity File No.:	01977	Cert. Date:	01/20/2022
Solution File No.:	02038	Lin. No.:	00036
Sequential File No.:	02038	Lin. Date:	01/20/2022
		Soln. No.:	00348
		Soln. Date:	05/21/2022
		File Date:	05/21/2022

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWL S3-0441
Control Solution %:	0.100%			Expires:	12/28/2023
Solution Control Lot:	21480			Bottle No.:	0248

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:34D		
Control 1 EC	0.099%	09:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:35D		
Control 2 EC	0.097%	09:36D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:37D		
Control 3 EC	0.098%	09:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: _____



Badge No.: 128

Date: 05/21/2022