

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE		
Calibration File No.:	02070	Calib. Date:	07/15/2022
Certification File No.:	02071	Cert. No.:	00046
Linearity File No.:	02072	Cert. Date:	07/15/2022
Solution File No.:	02093	Lin. No.:	00037
Sequential File No.:	02093	Lin. Date:	07/15/2022
		Soln. No.:	00354
		Soln. Date:	08/28/2022
		File Date:	08/28/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWL S3-0440
Solution Control Lot:	21480	Expires:	12/28/2023
		Bottle No.:	0503

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:24D		
Control 1 EC	0.099%	09:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:26D		
Control 2 EC	0.099%	09:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:27D		
Control 3 EC	0.100%	09:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: _____

Badge No.: 128

Date: 08/28/2022

